



LEGEND MAX[™]
ELISA Kit



Rat TNF- α

Cat. No. 438207

ELISA Kit for Accurate Quantitation of Rat TNF- α from Cell Culture Supernatant, Serum, Plasma and Other Biological Samples

BioLegend, Inc.
biolegend.com

It is highly recommended that this manual be read in its entirety before using this product. Do not use this kit beyond the expiration date.

For Research Purposes Only. Not for use in diagnostic or therapeutic procedures. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of BioLegend is strictly prohibited.



Table of Contents	Page
Introduction.....	2
Materials Provided.....	2
Materials to be Provided by the End-User.....	3
Storage Information.....	3
Health Hazard Warnings.....	3
Specimen Collection and Handling.....	4
Reagent and Sample Preparation.....	4
Assay Procedure.....	5
Assay Procedure Summary.....	7
Calculation of Results.....	8
Typical Data.....	8
Performance Characteristics.....	8
Specificity.....	8
Sensitivity.....	9
Recovery.....	9
Linearity.....	9
Intra-Assay Precision.....	9
Inter-Assay Precision.....	9
Biological Samples.....	9
Troubleshooting Guide.....	10
ELISA Plate Template.....	12

LEGEND MAX™ Rat TNF- α ELISA Kit

Introduction:

TNF- α (Tumor Necrosis Factor α), also termed Cachectin, Cytotoxic Factor (CF), Hemorrhagic Factor, Macrophage Cytotoxic Factor (MCF), is a potent multifunctional cytokine which can exert regulatory, inflammatory and cytotoxic effects on a wide range of normal lymphoid, non-lymphoid cells and tumor cells. TNF- α is secreted by macrophages, monocytes, neutrophils, T-cells, NK-cells, and many transformed cell lines. It exists as a soluble 17.5 kD protein and a 26 kD membrane-associated protein. The soluble protein form derives from the membrane form by proteolytic processing. Dephosphorylation of the membrane occurs by binding to soluble TNFRSF1A/ TNFR1. TNF- α binding to surface receptors elicits a wide array of biological activities including: cytolysis and cytostasis of many tumor cell lines in vitro, hemorrhagic necrosis of tumors in vivo, increased fibroblast proliferation and enhanced chemotaxis and phagocytosis in neutrophils. It has been shown that TNF- α plays a critical role in many physiological and pathological conditions, including neurodegenerative diseases and cancer.

BioLegend's LEGEND MAX™ Rat TNF- α ELISA Kit is a Sandwich Enzyme-Linked Immunosorbent Assay (ELISA) with a 96-well strip plate that is pre-coated with a monoclonal Armenian hamster anti-rat TNF- α capture antibody. The detection antibody is a biotinylated polyclonal goat anti-rat TNF- α antibody. This kit is specifically designed for the accurate quantitation of Rat TNF- α from cell culture supernatant, serum, plasma and other biological fluids. It is analytically validated with ready-to-use reagents.

Materials Provided:

Description	Quantity	Volume (per bottle)	Part #
Anti-Rat TNF- α Pre-coated 96-well Strip Microplate	1 plate		79501
Rat TNF- α Detection Antibody	1 bottle	12 mL	79521
Matrix A	1 vial	lyophilized	78303
Rat TNF- α Standard	1 vial	lyophilized	79554
Avidin-HRP D	1 bottle	12 mL	78237
Assay Buffer A	1 bottle	25 mL	78232
Wash Buffer (20X)	1 bottle	50 mL	78233
Substrate Solution F	1 bottle	12 mL	79132
Stop Solution	1 bottle	12 mL	79133
Plate Sealers	4 sheets		78101

Materials to be Provided by the End-User:

- Microplate reader able to measure absorbance at 450 nm
- Adjustable pipettes to measure volumes ranging from 1 μ L to 1,000 μ L
- Deionized water
- Wash bottle or automated microplate washer
- Log-Log graph paper or software for data analysis
- Tubes to prepare standard dilutions
- Timer
- Plate Shaker
- Polypropylene vials

Storage Information:

Store unopened kit components between 2°C and 8°C. Do not use this kit beyond its expiration date.

Opened or Reconstituted Components	
Microplate wells	If not all microplate strips are used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal. Store between 2°C and 8°C for up to one month.
Standard	The remaining reconstituted standard stock solution or Matrix A can be aliquoted into polypropylene vials and stored at -70°C for up to one month. Avoid repeated freeze-thaw cycles.
Matrix A	
Detection Antibody	Store opened reagents between 2°C and 8°C and use within one month.
Avidin-HRP D	
Assay Buffer A	
Wash Buffer (20X)	
Substrate Solution F	
Stop Solution	

Health Hazard Warnings:

1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin. Refer to the MSDS online at BioLegend’s website for details (www.biolegend.com/msds).
2. Substrate Solution F is harmful if inhaled or ingested. Avoid skin, eye and clothing contact.

LEGEND MAX™ Rat TNF- α ELISA Kit

3. To reduce the likelihood of blood-borne transmission of infectious agents, handle all serum, plasma and other biological fluids in accordance with NCCLS regulations.
4. Stop Solution contains strong acid. *Wear eye, hand, and face protection.*
5. Before disposing the plate, rinse with an excess amount of tap water.

Specimen Collection and Handling:

Specimens should be clear and non-hemolyzed. If possible, unknown samples should be run at a number of dilutions to determine the optimal dilution factor that will ensure accurate quantitation.

Cell Culture Supernatant: If necessary, centrifuge all samples to remove debris prior to analysis. It is recommended that samples be stored at $< -70^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum: Use a serum separator tube and allow clotting for at least 30 minutes, then centrifuge for 10 minutes at $1,000 \times g$. Remove serum layer and assay immediately or store serum samples at $< -70^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma: Collect blood samples in citrate, heparin or EDTA containing tubes. Centrifuge for 10 minutes at $1,000 \times g$ within 30 minutes of collection. Assay immediately or store plasma samples at $< -70^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Reagent and Sample Preparation:

Note: All reagents should be diluted immediately prior to use.


1. Dilute the 20X Wash Buffer to 1X with deionized water. For example, make 1 liter of 1X Wash Buffer by adding 50 mL of 20X Wash Buffer to 950 mL of deionized water. If crystals have formed in the 20X Wash Buffer, bring to room temperature and vortex until dissolved.
2. Reconstitute the lyophilized Rat TNF- α Standard by adding the volume of Assay Buffer A to make the 20 ng/mL standard stock solution (Refer to LEGEND MAX Kit Lot-Specific Certificate of Analysis/LEGEND MAX Kit Protocol). Allow the reconstituted standard to sit at room temperature for 15 minutes, then briefly vortex to mix completely.
3. If serum or plasma samples will be assayed, reconstitute the lyophilized Matrix A by dispensing 2 mL of deionized water into the vial and allow the reconstituted Matrix A to sit at room temperature for 15 minutes, then vortex to mix completely.

- In general, cell culture supernatant samples are analyzed without dilutions. However, if dilutions are required, use the control culture medium or Assay Buffer A as the sample diluent.
- For measuring serum or plasma samples, a 2-fold dilution of the samples with Assay Buffer A is suggested, e.g. add 50 μ L of sample to 50 μ L of Assay Buffer A. If further dilution is required samples should be diluted with Matrix A.

Assay Procedure:

Note: Do not mix reagents from different kits or lots. Reagents from different manufacturers should not be used with this kit.

- Bring all reagents to room temperature prior to use. It is strongly recommended that all standards and samples be run in duplicate or triplicate. A standard curve is required for each assay.
- If not all microplate strips will be used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal.
- Prepare 500 μ L of the 500 pg/mL top standard by diluting 12.5 μ L of the standard stock solution in 487.5 μ L of Assay Buffer A. Perform six two-fold serial dilutions of the 500 pg/mL top standard in separate tubes using Assay Buffer A as the diluent. Thus, the Rat TNF- α standard concentrations in the tubes are 500 pg/mL, 250 pg/mL, 125 pg/mL, 62.5 pg/mL, 31.3 pg/mL, 15.6 pg/mL, and 7.8 pg/mL, respectively. Assay Buffer A serves as the zero standard (0 pg/mL).



- Wash the plate 4 times with at least 300 μ L of 1X Wash Buffer per well and blot any residual buffer by firmly tapping the plate upside down on absorbent paper. All subsequent washes should be performed similarly.

LEGEND MAX™ Rat TNF- α ELISA Kit

5. For measuring samples of cell culture supernatant:

- a) Add 50 μ L of Assay Buffer A to each well that will contain either standard dilutions or samples.
- b) Add 50 μ L of standard dilutions or samples to the appropriate wells.


For measuring serum or plasma samples:

- a) Add 50 μ L of Matrix A to each well that will contain the standard dilutions. Add 50 μ L of Assay Buffer A to each well that will contain samples.
 - b) Add 50 μ L of standard dilutions to the wells containing Matrix A. Add 50 μ L of diluted serum or plasma samples to the wells containing Assay Buffer A.
6. Seal the plate with a Plate Sealer included in the kit and incubate the plate at room temperature for 2 hours while shaking at 200 rpm.
 7. Discard the plate contents into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
 8. Add 100 μ L of Rat TNF- α Detection Antibody solution to each well, seal the plate and incubate at room temperature for 1 hour while shaking.
 9. Discard the contents of the plate into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
 10. Add 100 μ L of Avidin-HRP D solution to each well, seal the plate and incubate at room temperature for 30 minutes while shaking.
 11. Discard the contents of the plate into a sink, then wash the plate 5 times with 1X Wash Buffer as in step 4. For this final wash, soak wells in 1X Wash Buffer for 30 seconds to 1 minute for each wash. This will help minimize background.
 12. Add 100 μ L of Substrate Solution F to each well and incubate for 15 minutes in the dark. Wells containing Rat TNF- α in samples should turn blue in color with an intensity proportional to its concentration. It is not necessary to seal the plate during this step.
 13. Stop the reaction by adding 100 μ L of Stop Solution to each well. The solution color should change from blue to yellow.
 14. Read absorbance at 450 nm within 30 minutes. If the reader is capable of reading at 570 nm, the absorbance at 570 nm can be subtracted from the absorbance at 450 nm.

Assay Procedure Summary


- For measuring cell culture supernatant:**


Wash 4 times
Add 50 μ L Assay Buffer A to standard and sample wells




For measuring serum or plasma:


Wash 4 times
Add 50 μ L Matrix A to standard wells
Add 50 μ L Assay Buffer A to sample wells



- Add 50 μ L diluted standards to standard wells
Add 50 μ L samples to sample wells
Incubate 2 hr, RT, shaking





Add 50 μ L diluted standards to standard wells
Add 50 μ L diluted samples to sample wells
Incubate 2 hr, RT, shaking


- Wash 4 times
Add 100 μ L Detection Antibody solution
Incubate 1 hr, RT, shaking


- Wash 4 times
Add 100 μ L Avidin-HRP D solution
Incubate 30 min, RT, shaking


- Wash 5 times
Add 100 μ L Substrate Solution F
Incubate 15 min, RT, in the dark


- Add 100 μ L Stop Solution


- Read absorbance at 450 nm and 570 nm

LEGEND MAX™ Rat TNF- α ELISA Kit

Calculation of Results:


The data can be best calculated with computer-based curve-fitting software using a 5- or 4-parameter logistics curve-fitting algorithm. If an appropriate software is not available, use log-log graph paper to determine sample concentrations. Determine the mean absorbance for each set of duplicate or triplicate standards, controls, and samples. Plot the standard curve on log-log graph paper with cytokine concentration on the X-axis and absorbance on the Y-axis. Draw a best fit line through the standard points. To determine the unknown cytokine concentrations, find the mean absorbance value of the unknown concentration on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the cytokine concentration.

Because serum/plasma samples have been diluted prior to the assay, their measured concentrations must be multiplied by the dilution factor.

If cell culture supernatant samples were diluted, multiply the concentration by the appropriate dilution factor. If a test sample's absorbance value falls outside the linear portion of the standard curve, the test sample needs to be re-analyzed at a higher (or lower) dilution as appropriate.

Typical Data:

This standard curve was generated at BioLegend for demonstration purposes only. A standard curve must be run with each assay.



Performance Characteristics:

Specificity: This kit recognizes mouse TNF-α and there is 0.04% cross reactivity with human TNF-α. No other cross reactivity was observed when this kit analyzed the following cytokines/chemokines, each at a concentration of 50 ng/mL

Rat	IL-1α, IL-1β, IL-2, IL-4, IL-5, IL-6, IL-9, IL-10, IL-12 (p70), IL-13, IL-17A, IL-17E, IL-17F, IL-18, IL-22, IL-23, IL-33, CCL20, GM-CSF, IFN-γ, KC, MCP-1, RANTES
-----	--------------------------------------------------------------------------------------------------------------------------------------------------------------------

Sensitivity: The minimum detectable concentration of Rat TNF-α is 4.2 pg/mL.

Recovery: TNF-α at concentrations of 500 pg/mL, 125 pg/mL, and 31.25 pg/mL was spiked into 4 strains of rat serum samples, and then analyzed with the LEGEND MAX™ Rat TNF-α ELISA Kit. On average, 107.1% of the cytokine was recovered from the serum samples.

Linearity: Serum samples representing four strains of rat containing high concentrations of TNF-α were diluted with matrix A to produce samples with values within the dynamic range and then assayed with the kit to determine the dilutional linearity. On average, 110.7 % of the expected cytokine was detected from the serum samples.

Intra-Assay Precision: Twelve replicates each of two different TNF-α concentrations were tested on one plate.

Concentration	Sample 1	Sample 2
Number of Replicates	12	12
Mean Concentration (pg/mL)	170.8	63.5
Standard Deviation	14.1	4.7
% CV	8.3	7.4

Inter-Assay Precision: Two concentrations of TNF-α , each with sixteen replicates, were assayed on four different plates.

Concentration	Sample 1	Sample 2
Number of Assays	4	4
Mean Concentration (pg/mL)	156.2	52.9
Standard Deviation	4.6	5.6
% CV	3.0	10.5

Biological Samples: Rat spleenocytes (1×10^6 cells/mL) were stimulated with 1 μg/mL of coated anti-human CD3 and 0.5 μg/mL of soluble anti-human CD28 antibodies at 37°C for 48 hours. The cell culture supernatants were collected and assayed for the concentration of biologically-derived rat TNF-α. The detected concentrations averaged from 485 pg/mL in stimulated samples and was undetectable in unstimulated samples.

Normal Serum Concentration - Four pooled rat serum samples, each from a different strain, were tested for basal levels of rat TNF-α. None of them showed detectable levels of TNF-α.

LEGEND MAX™ Rat TNF- α ELISA Kit

Troubleshooting Guide:

Problem	Probable Cause	Solution
High Background	Background wells were contaminated	Avoid cross-well contamination by using the provided plate sealers. Use multichannel pipettes and change tips between pipetting samples and reagents.
	Insufficient washes	Increase number of washes. Increase soaking time between washes prior to addition of substrate solution.
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells. Use a clean container prior to pipetting substrate solution into wells.
No or poor signal	Detection Antibody, Avidin-HRP or Substrate solution were NOT added	Rerun the assay and follow the protocol.
	Wrong reagent or reagents were added in wrong sequential order	
	Insufficient plate agitation	The plate should be agitated during all incubation steps using a plate shaker at a speed where solutions in wells are within constant motion without splashing.
	The wash buffer contains Sodium Azide (NaN ₃)	Avoid Sodium Azide contamination in the wash buffer as it inhibits HRP activity.
	Incubations were done at an inappropriate temperature, timing or without agitation	Rerun the assay and follow the protocol.
Low or poor standard curve signal	The standard was incorrectly reconstituted or diluted	Adjust the calculations and follow the protocol.
	Standard was inappropriately stored	Store the reconstituted standard stock solution in polypropylene vials at -70°C. Avoid repeated freeze-thaw cycles.
	Reagents added to wells with incorrect concentrations	Check for pipetting errors and the correct reagent volume.

Problem	Probable Cause	Solution
Signal is high, standard curves have saturated signal	Standard reconstituted with less volume than required	Reconstitute new lyophilized standard with the correct volume of solution recommended in the protocol.
	Standards/samples, detection antibody, Avidin-HRP or substrate solution were incubated for too long	Rerun the assay and follow the protocol.
Sample readings are out of range	Samples contain no or below detectable levels of the analyte	If samples are below detectable levels, it may be possible to use a larger sample volume. Contact technical support for appropriate protocol modifications.
	Samples contain analyte concentrations greater than highest standard point	Samples may require dilution and analysis.
High variation in samples and/or standards	Multichannel pipette errors	Confirm that pipette calibrations are accurate.
	Plate washing was not adequate or uniform	Ensure pipette tips are tightly secured. Ensure uniformity in all wash steps.
	Non-homogenous samples	Thoroughly mix samples before assaying.
	Samples may have high particulate matter	Remove particulate matter by centrifugation.
	Cross-well contamination	Do not reuse plate sealers. Always change tips for reagent additions. Ensure that pipette tips do not touch the reagents on the plate.

ELISA Plate Template												
	1	2	3	4	5	6	7	8	9	10	11	12
A												
B												
C												
D												
E												
F												
G												
H												



LEGEND MAX™ Kits are manufactured by **BioLegend Inc.**

8999 BioLegend Way

San Diego, CA 92121

Tel: 1.858.768.5800

Tel US & Canada Toll-Free: 1.877.Bio-Legend (1.877.246.5343)

Fax: 1.877.455.9587

Email: info@biolegend.com

biolegend.com